



File E149685

Vol 1

Issued: 1992-08-17
Revised: 2009-04-13

FOLLOW-UP SERVICE PROCEDURE
(TYPE R)

COMPONENT - POWER SUPPLIES, MEDICAL AND DENTAL
(QQHM2)

Manufacturer: ELPAC ELECTRONICS (SHENZHEN) INC
(940730-001) MINLIDA IND GARDEN, SAST BLDG,3RD FL S
HONGHUALING IND ZONE
NANSHAN, XILI TOWN
SHENZHEN,
GUANGDONG 518055 CHINA

Applicant: INTERNATIONAL COMPONENTS CORP-DESIGN
(100408-596) CENTER
3RD FL, WESTERN CHANGHUA BLDG
929 XING YE RD, NANCUN TOWN
PANYU
GUANGZHOU 511442 GUANGDONG CHINA

Recognized Company: INTERNATIONAL COMPONENTS CORP (E329536)
(171557-001) SUITE 900
4 WESTBROOK CORPORATE CENTER
WESTCHESTER IL 60154

This Procedure authorizes the above manufacturer to use the marking specified by Underwriters Laboratories Inc.(UL), or any authorized licensee of UL, only on products covered by this Procedure, in accordance with the applicable UL Services Agreement.

The prescribed Mark or Marking shall be used only at the above manufacturing location on such products which comply with this Procedure and any other applicable requirements.

The Procedure contains information for the use of the above named Manufacturer and representatives of Underwriters Laboratories Inc. and is not to be used for any other purpose. It is lent to the Manufacturer with the understanding that it is not to be copied, either wholly or in part, and that it will be returned to Underwriters Laboratories Inc. (UL) or any authorized licensee of UL, upon request.

This PROCEDURE, and any subsequent revision, is the property of Underwriters Laboratories Inc.(UL) and the authorized licensee of UL and is not transferable.

Underwriters Laboratories Inc.

Stephen Hewson
Senior Vice President
Global Follow-Up Service Operations

William R. Carney
Director
North American Certification Program

COVER PAGE FOR TEST REPORT

Product Category:	Power Supplies, Medical and Dental		
Product Category CCN:	QQHM2, QQHM8		
Test Procedure:	Component Recognition		
Product:	Switch Mode Power Supply		
Model/Type Reference:	MW24 Series: MW2412, MW2415, MW2418, and MW2424		
Rating(s):	Model:	Input:	Output (dc):
	MW2412	100-240 V ac, 0.7 A, 47-63 Hz	12 V, 2.0 A
	MW2415	100-240 V ac, 0.7 A, 47-63 Hz	15 V, 1.6 A
	MW2418	100-240 V ac, 0.7 A, 47-63 Hz	18 V, 1.33 A
	MW2424	100-240 V ac, 0.7 A, 47-63 Hz	24 V, 1.0 A
Standards:	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)		
Applicant Name and Address:	ELPAC ELECTRONICS INC 1562 REYNOLDS AVE IRVINE CA 92614 UNITED STATES		
This Report includes the following parts, in addition to this cover page:			
1. Specific Inspection Criteria			
2. Specific Technical Criteria			
3. Clause Verdicts			
4. Critical Components			
5. Test Results			
6. National Differences			
7. Enclosures			

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Test Report By:



Linus Park
Lead Engineering Associate
Underwriters Laboratories Inc.

Reviewed By:



Marc M. Mouser
Manager
Underwriters Laboratories Inc.