



Ref. Certif. No.

US/9356/UL

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switch Mode Power Supply

Name and address of the applicant
Nom et adresse du demandeur

Elpac Electronics Inc.
Elpac Power Systems, Div of
1562 Reynolds Ave
Irvine, CA 92614, USA

Name and address of the manufacturer
Nom et adresse du fabricant

Elpac Electronics Inc.
Elpac Power Systems, Div of
1562 Reynolds Ave
Irvine, CA 92614, USA

Name and address of the factory
Nom et adresse de l'usine

Elpac Electronics (Shenzhen) Inc.
Minlida Ind Gardner, Sast Bldg, 5th Fl S
Honghualing Industrial Zone
Shenzhen, Nanshan, Xili Town, Guangdong 518055, China

Rating and principal characteristics
Valeurs nominales et caractéristiques principales

Models MSM40A, MSM40B, MSM40D, Input: 100-240 V~, 50/60 Hz, 1.3 A
Model MSM40 A, Output: +5.1 Vdc, 4 A; +12 Vdc, 2 A; -12 Vdc, 0.4 A
Model MSM40B, Output: +5.1 Vdc, 4 A; +15 Vdc, 2 A; -15 Vdc, 0.4 A
Model MSM40D, Output: +5.1 Vdc, 4 A; +24 Vdc, 1 A; -12 Vdc, 0.4 A
Total Maximum 40 W

mark (if any)
marque de fabrique (si elle existe)



Model / Type Ref.
Ref. de type

MSM40A, MSM40B, MSM40D

Additional information (if necessary)
Information complémentaire (si nécessaire)

The CB Test Report comprises 4 Enclosures.

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

PUBLICATION

EDITION

**IEC 60601-1 (1988) Second Edition,
with Amendment No. 1 (1991) and No. 2 (1995) with the exception of:
Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and
Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common
Modifications. See Test Report for National Differences.**

as shown in the Test Report Ref. No.
which forms part of this Certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue partie de ce Certificat

E149685-A12-CB-1

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



**Underwriters
Laboratories Inc.®**

Underwriters Laboratories Inc. / Certification Programs Office
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* 1-847-272-8800, Ext. 43008 FAX INT* 1-847-272-9562
email: jolanta.m.wroblewska@us.ul.com

Date:
Issued: 2005 May 3

Signature:

Jolanta M. Wroblewska